

HPLC Application

ID No.: 13625

Diperodon

Column: N/A

Dimensions: N/A

Order No: N/A

Elution Type: N/A

Eluent A: Hexane/Dichloroethane/Ethanol-TFA 60:35:5 ethanol-TFA 20:1

Gradient

Step No.	Time (min)	Pct A
----------	------------	-------


Profile:

1	0	100
---	---	-----

Flow Rate: 1 mL/min

Col. Temp.: ambient

Detection: UV-Vis Abs.-Variable Wave.(UV) @ 242 nm (ambient)



ANALYTES:

1 Diperodon

Retention Time: 18.3 min

2 Diperodon

Retention Time: 18.85 min

